

i-STAT Calibration Verification Set B0797211

LOT L1 B07339, L2 B07347, L3 B07352, L4 B07361, L5 B07341

Exp.: 2009-06

CLEW: A17/C17/H17/J17

CG8+, EG7+, EG6+ & CG4+		LOT G, H, J, L, N, U, W or Y				
		1	2	3	4	5
Na	mmol/L	102.6	120.1	134.7	159.8	177.4
	mEq/L	99.2-106.0	116.7-123.5	130.8-138.6	155.4-164.2	172.2-182.6
K	mmol/L	2.11	2.88	3.70	6.22	7.69
	mEq/L	1.85-2.37	2.62-3.14	3.41-3.99	5.83-6.61	7.19-8.19
pH		6.820	7.175	7.455	7.696	8.156
		6.771-6.869	7.126-7.224	7.406-7.504	7.647-7.745	8.107-8.205
PCO ₂	mmHg	88.27	60.30	25.82	18.77	15.73
		79.62-96.92	54.18-66.42	20.43-31.21	13.63-23.91	10.83-20.63
	kPa	11.766	8.038	3.442	2.502	2.097
		10.614-12.919	7.222-8.854	2.723-4.160	1.816-3.188	1.444-2.750
PO ₂	mmHg	69.2	81.6	107.3	134.9	323.2
		56.2-82.2	68.6-94.6	93.4-121.2	117.4-152.4	271.4-375.0
	kPa	9.22	10.88	14.30	17.98	43.08
		7.50-10.95	9.15-12.61	12.45-16.16	15.65-20.31	36.17-49.99
iCa	mmol/L	2.257	1.437	1.183	0.756	0.317
		2.127-2.387	1.348-1.526	1.109-1.257	0.691-0.821	0.260-0.374
	mg/dL	9.03	5.75	4.73	3.02	1.27
		8.51-9.55	5.39-6.11	4.44-5.03	2.76-3.28	1.04-1.50
	mEq/L	4.51	2.87	2.37	1.51	0.63
4.25-4.77		2.70-3.05	2.22-2.51	1.38-1.64	0.52-0.75	
Glucose/ Glu	mg/dL	28.7	41.7	122.4	281.6	600.2
		19.5-37.8	34.9-48.5	109.9-134.9	252.8-310.3	532.1-668.4
	g/L	0.287	0.417	1.224	2.816	6.002
		0.195-0.378	0.349-0.485	1.099-1.349	2.528-3.103	5.321-6.684
	mmol/L	1.59	2.32	6.80	15.63	33.32
1.08-2.10		1.94-2.69	6.10-7.49	14.03-17.23	29.53-37.10	
Lactate/ Lac	mmol/L	17.470	7.884	2.104	0.830	0.442
		14.90-20.04	7.01-8.75	1.70-2.51	0.42-1.24	0.03-0.85
	mg/dL	157.4	71.0	18.96	7.48	3.98
		134.3-180.5	63.2-78.9	15.3-22.6	3.8-11.2	0.3-7.7
	g/L	1.574	0.710	0.1896	0.0748	0.0398
1.343-1.805		0.632-0.789	0.153-0.226	0.038-0.112	0.003-0.077	

i-STAT Calibration Verification Set B0797211

LOT L1 B07339, L2 B07347, L3 B07352, L4 B07361, L5 B07341

Exp.: 2009-06

CLEW: A17/C17/H17/J17

CG8+, EG7+, EG6+, G3+, Crea & CG4+		LOT A, B, C, D, E, F, K, M, P, R, S or T				
		1	2	3	4	5
Na	mmol/L	102.2	121.1	136.4	162.0	180.6
	mEq/L	98.8-105.6	117.7-124.5	132.5-140.3	157.6-166.4	175.4-185.8
K	mmol/L	2.09	2.90	3.73	6.31	7.84
	mEq/L	1.83-2.35	2.64-3.16	3.44-4.02	5.92-6.70	7.34-8.34
pH		6.833	7.177	7.453	7.685	8.129
		6.784-6.882	7.128-7.226	7.404-7.502	7.636-7.734	8.080-8.178
PCO ₂	mmHg	90.86	61.99	26.46	19.30	16.86
		81.96-99.76	55.87-68.11	21.07-31.85	14.16-24.44	11.96-21.76
	kPa	12.112	8.263	3.527	2.573	2.247
		10.925-13.298	7.447-9.080	2.809-4.245	1.887-3.258	1.594-2.900
PO ₂	mmHg	74.8	86.8	117.8	147.5	355.4
		61.8-87.8	73.8-99.8	102.5-133.1	128.4-166.6	303.6-407.2
	kPa	9.97	11.57	15.70	19.66	47.37
		8.24-11.70	9.84-13.30	13.67-17.74	17.11-22.21	40.46-54.29
iCa	mmol/L	2.258	1.468	1.215	0.777	0.326
		2.128-2.388	1.377-1.559	1.139-1.291	0.712-0.842	0.269-0.383
	mg/dL	9.03	5.87	4.86	3.11	1.30
		8.51-9.55	5.51-6.24	4.56-5.16	2.85-3.37	1.08-1.53
	mEq/L	4.52	2.94	2.43	1.55	0.65
4.26-4.78		2.75-3.12	2.28-2.58	1.42-1.68	0.54-0.77	
Glucose/ Glu	mg/dL	28.7	41.7	122.4	281.6	600.2
		19.5-37.8	34.9-48.5	109.9-134.9	252.8-310.3	532.1-668.4
	g/L	0.287	0.417	1.224	2.816	6.002
		0.195-0.378	0.349-0.485	1.099-1.349	2.528-3.103	5.321-6.684
	mmol/L	1.59	2.32	6.80	15.63	33.32
1.08-2.10		1.94-2.69	6.10-7.49	14.03-17.23	29.53-37.10	
Creatinine/ Crea	mg/dL	14.57	4.31	1.85	0.68	0.49
		11.96-17.18	3.77-4.85	1.49-2.21	0.32-1.04	0.13-0.85
	umol/L	1288.0	381.0	163.5	60.1	43.3
		1057.4-1518.6	333.2-428.8	131.9-195.2	28.5-91.8	11.7-75.0
Lactate/ Lac	mmol/L	17.470	7.884	2.104	0.830	0.442
		14.90-20.04	7.01-8.75	1.70-2.51	0.42-1.24	0.03-0.85
	mg/dL	157.4	71.0	18.96	7.48	3.98
		134.3-180.5	63.2-78.9	15.3-22.6	3.8-11.2	0.3-7.7
	g/L	1.574	0.710	0.1896	0.0748	0.0398
1.343-1.805		0.632-0.789	0.153-0.226	0.038-0.112	0.003-0.077	



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022113-01 01
722113-01A
2-Apr-2009



i-STAT Calibration Verification Set B0797211

LOT L1 B07339, L2 B07347, L3 B07352, L4 B07361, L5 B07341

Exp.: 2009-06

CLEW: A17/C17/H17/J17

EC8+, 6+, EC4+, E3+ & G		LOT A, B, C, D, E or F				
		1	2	3	4	5
Na	mmol/L	102.6	119.9	134.4	159.2	176.8
	mEq/L	99.2-106.0	116.5-123.3	130.5-138.3	154.8-163.6	171.6-182.0
K	mmol/L	2.10	2.85	3.65	6.12	7.55
	mEq/L	1.84-2.36	2.59-3.11	3.36-3.94	5.73-6.51	7.05-8.05
Cl	mmol/L	67.7	78.6	92.4	115.5	126.3
	mEq/L	63.7-71.7	74.3-82.9	87.9-96.9	109.5-121.5	118.8-133.8
pH		6.831	7.179	7.456	7.702	8.163
		6.782-6.880	7.130-7.228	7.407-7.505	7.653-7.751	8.114-8.212
PCO₂	mmHg	89.24	59.87	24.29	17.24	13.65
		80.50-97.98	53.75-65.99	18.90-29.68	12.10-22.38	8.75-18.55
	kPa	11.896	7.981	3.238	2.298	1.820
		10.730-13.061	7.164-8.797	2.520-3.956	1.612-2.984	1.167-2.473
Glucose/ Glu	mg/dL	28.7	41.7	122.4	281.6	600.2
		19.5-37.8	34.9-48.5	109.9-134.9	252.8-310.3	532.1-668.4
	g/L	0.287	0.417	1.224	2.816	6.002
		0.195-0.378	0.349-0.485	1.099-1.349	2.528-3.103	5.321-6.684
	mmol/L	1.59	2.32	6.80	15.63	33.32
		1.08-2.10	1.94-2.69	6.10-7.49	14.03-17.23	29.53-37.10
BUN	mg/dL	131.7	52.9	9.8	5.0	2.1
		116.1-147.3	46.6-59.2	7.2-12.4	2.9-7.1	0.0-4.2
Urea	mmol/L	47.02	18.89	3.50	1.79	0.75
		41.46-52.57	16.65-21.12	2.56-4.44	1.03-2.54	0.00-1.50
	mg/dL	281.8	113.2	21.0	10.7	4.5
		248.5-315.2	99.8-126.6	15.4-26.6	6.2-15.2	0.0-9.0
	g/L	2.818	1.132	0.210	0.107	0.045
		2.485-3.152	0.998-1.266	0.154-0.266	0.062-0.152	0.000-0.090

IVD



CE



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022113-01 01
722113-01A
2-Apr-2009

