

i-STAT BNP Calibration Verification Control Set

LOT M08343 (L1 M083431, L2 M083432, L3 M083433)

 Exp.: 2009-06

CLEW: A17

BNP **LOT** **B**

		L1	L2	L3
BNP	pg/mL	99	948	2927
		54-144	607-1289	1873-3981

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022160-01 01
722160-01A
3-Apr-2009



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

