


i-STAT PT Level 2 Control

LOT B08177

 Exp.: 2009-05


CLEW: A17/C17/J17/H17

PT **LOT** G,J,M or N

| | \bar{x} (Mean) | R (Range) |
|----------|------------------|-----------|
| PT - INR | 2.65 | 2.0-3.3 |
| PT - sec | 30.4 | 22.7-38.0 |

IVD



 Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022181-01 01
722181-01A
3-Apr-2009

