

i-STAT Calibration Verification Set B0953406

LOT L1 B09155, L2 B09153, L3 B09154, L4 B09160, L5 B09156

Exp.: 2010-11

CLEW: A18/C18/H18/J18

CG4+	LOT					J or W				
	1	2	3	4	5					
pH	6.828	7.168	7.455	7.704	8.134					
	6.779-6.877	7.119-7.217	7.406-7.504	7.655-7.753	8.085-8.183					
PCO ₂	mmHg	90.08	61.19	26.09	18.16	16.67				
		81.25-98.91	55.07-67.31	20.70-31.48	13.02-23.30	11.77-21.57				
	kPa	12.008	8.157	3.478	2.421	2.222				
PO ₂	mmHg	70.2	81.4	108.3	136.4	324.2				
		57.2-83.2	68.4-94.4	94.3-122.3	118.7-154.1	272.4-376.0				
	kPa	9.36	10.85	14.44	18.18	43.22				
Lactate/ Lac	mmol/L	17.206	7.729	2.020	0.786	0.434				
		14.68-19.73	6.88-8.58	1.61-2.43	0.38-1.19	0.03-0.84				
	mg/dL	155.0	69.6	18.20	7.08	3.91				
		132.2-177.8	62.0-77.3	14.5-21.9	3.4-10.8	0.2-7.6				
	g/L	1.550	0.696	0.1820	0.0708	0.0391				
		1.322-1.778	0.620-0.773	0.145-0.219	0.034-0.108	0.002-0.076				

i-STAT Calibration Verification Set B0953406

LOT L1 B09155, L2 B09153, L3 B09154, L4 B09160, L5 B09156

Exp.: 2010-11

CLEW: A18/C18/H18/J18

CG8+, EG7+, EG6+, G3+, Crea & CG4+	LOT					A, B, C, D, E, F, K, M, P, R, S or T				
	1	2	3	4	5					
Na	mmol/L	101.6	120.3	135.6	161.2	180.6				
	mEq/L	98.2-105.0	116.9-123.7	131.7-139.5	156.8-165.6	175.4-185.8				
K	mmol/L	2.07	2.86	3.71	6.27	7.84				
	mEq/L	1.81-2.33	2.60-3.12	3.42-4.00	5.88-6.66	7.34-8.34				
pH		6.837	7.169	7.451	7.690	8.109				
		6.788-6.886	7.120-7.218	7.402-7.500	7.641-7.739	8.060-8.158				
PCO ₂	mmHg	91.59	61.42	26.86	19.55	18.35				
		82.62-100.56	55.30-67.54	21.47-32.25	14.41-24.69	13.45-23.25				
	kPa	12.209	8.187	3.580	2.606	2.446				
PO ₂	mmHg	74.0	86.2	114.8	144.5	345.8				
		61.0-87.0	73.2-99.2	99.9-129.7	125.8-163.2	294.0-397.6				
	kPa	9.86	11.49	15.30	19.26	46.10				
iCa	mmol/L	2.191	1.454	1.206	0.775	0.356				
		2.061-2.321	1.364-1.544	1.131-1.281	0.710-0.840	0.299-0.413				
	mg/dL	8.76	5.82	4.82	3.10	1.42				
Glucose/ Glu	mg/dL	8.25-9.28	5.45-6.18	4.52-5.12	2.84-3.36	1.20-1.65				
		4.38	2.91	2.41	1.55	0.71				
	mEq/L	4.12-4.64	2.73-3.09	2.26-2.56	1.42-1.68	0.60-0.83				
Creatinine/ Crea	mg/dL	28.5	41.3	122.4	281.0	600.5				
		19.3-37.7	34.5-48.1	109.9-134.9	252.3-309.7	532.3-668.7				
	g/L	0.285	0.413	1.224	2.810	6.005				
Lactate/ Lac	mmol/L	0.193-0.377	0.345-0.481	1.099-1.349	2.523-3.097	5.323-6.687				
		1.58	2.29	6.79	15.60	33.33				
		1.07-2.09	1.91-2.67	6.10-7.49	14.00-17.19	29.55-37.12				
Creatinine/ Crea	mg/dL	14.81	4.30	1.84	0.64	0.44				
		12.16-17.46	3.76-4.84	1.48-2.20	0.28-1.00	0.08-0.80				
	umol/L	1309.2	380.1	162.7	56.6	38.9				
Lactate/ Lac	mmol/L	1074.8-1543.6	332.5-427.8	131.0-194.3	24.9-88.2	7.2-70.6				
		17.888	7.940	2.168	0.847	0.446				
	mg/dL	15.26-20.52	7.06-8.82	1.76-2.58	0.44-1.26	0.04-0.85				
Lactate/ Lac	mg/dL	161.2	71.5	19.53	7.63	4.02				
		137.5-184.9	63.7-79.4	15.9-23.2	4.0-11.3	0.3-7.7				
	g/L	1.612	0.715	0.1953	0.0763	0.0402				
		1.375-1.849	0.637-0.794	0.159-0.232	0.040-0.113	0.003-0.077				



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

022233-01 01
722233-01B
30-Oct-2009



i-STAT Calibration Verification Set B0953406

LOT L1 B09155, L2 B09153, L3 B09154, L4 B09160, L5 B09156

Exp.: 2010-11

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+, E3+ & G		LOT A, B, C, D, E or F				
		1	2	3	4	5
Na	mmol/L	102.4	119.6	134.1	158.8	176.9
	mEq/L	99.0-105.8	116.2-123.0	130.2-138.0	154.4-163.2	171.7-182.1
K	mmol/L	2.09	2.84	3.65	6.14	7.65
	mEq/L	1.83-2.35	2.58-3.10	3.36-3.94	5.75-6.53	7.15-8.15
Cl	mmol/L	65.9	78.1	92.3	114.9	126.1
	mEq/L	61.9-69.9	73.8-82.4	87.8-96.8	108.9-120.9	118.6-133.6
pH		6.837	7.174	7.457	7.709	8.141
		6.788-6.886	7.125-7.223	7.408-7.506	7.660-7.758	8.092-8.190
PCO₂	mmHg	90.02	59.55	24.30	16.66	14.62
		81.20-98.84	53.43-65.67	18.91-29.69	11.52-21.80	9.72-19.52
	kPa	12.000	7.938	3.239	2.221	1.949
		10.824-13.175	7.122-8.754	2.521-3.958	1.535-2.906	1.296-2.602
Glucose/ Glu	mg/dL	28.5	41.3	122.4	281.0	600.5
		19.3-37.7	34.5-48.1	109.9-134.9	252.3-309.7	532.3-668.7
	g/L	0.285	0.413	1.224	2.810	6.005
		0.193-0.377	0.345-0.481	1.099-1.349	2.523-3.097	5.323-6.687
	mmol/L	1.58	2.29	6.79	15.60	33.33
		1.07-2.09	1.91-2.67	6.10-7.49	14.00-17.19	29.55-37.12
BUN	mg/dL	133.9	53.3	10.1	5.1	2.1
		118.1-149.7	47.0-59.6	7.5-12.7	3.0-7.2	0.0-4.2
Urea	mmol/L	47.80	19.03	3.61	1.82	0.75
		42.15-53.45	16.78-21.28	2.67-4.54	1.07-2.57	0.00-1.50
	mg/dL	286.5	114.1	21.6	10.9	4.5
		252.7-320.4	100.6-127.5	16.0-27.2	6.4-15.4	0.0-9.0
	g/L	2.865	1.141	0.216	0.109	0.045
		2.527-3.204	1.006-1.275	0.160-0.272	0.064-0.154	0.000-0.090



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022233-01 01
722233-01B
30-Oct-2009

