

i-STAT Level 1 Control

LOT B09153

Exp.: 2010-11

CLEW: A18/C18/H18/J18

CG4+		LOT	J or W	
			\bar{x} (Mean)	R (Range)
pH			7.168	7.118-7.218
PCO ₂	mm Hg		61.2	53.2-69.2
	kPa		8.16	7.09-9.22
PO ₂	mm Hg		81	66-96
	kPa		10.9	8.9-12.9
Lactate/Lac	mmol/L		7.73	6.81-8.65
	mg/dL		69.6	61.3-77.9
	g/L		0.696	0.613-0.779

CG8+, EG7+, EG6+, G3+, Crea & CG4+		LOT	A, B, C, D, E, F, K, M, P, R, S or T	
			\bar{x} (Mean)	R (Range)
Na	mmol/L, mEq/L		120	116-124
K	mmol/L, mEq/L		2.9	2.5-3.3
iCa	mmol/L		1.45	1.34-1.56
	mg/dL		5.8	5.4-6.3
	mEq/L		2.9	2.7-3.1
pH			7.169	7.119-7.219
PCO ₂	mm Hg		61.4	53.4-69.4
	kPa		8.19	7.12-9.25
PO ₂	mm Hg		86	71-101
	kPa		11.5	9.5-13.5
Glucose/Glu	mg/dL		41	32-50
	g/L		0.41	0.32-0.50
	mmol/L		2.3	1.9-2.7
Creatinine/Crea	mg/dL		4.3	3.5-5.1
	umol/L		380	309-451
Lactate/Lac	mmol/L		7.94	7.02-8.86
	mg/dL		71.5	63.3-79.8
	g/L		0.715	0.633-0.798

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299


022234-01 01
722234-01B
30-Oct-2009



Abbott

i-STAT Level 1 Control

LOT B09153

 Exp.: 2010-11

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+, E3+ & G		LOT	A, B, C, D, E or F
		\bar{x} (Mean)	R (Range)
Na	mmol/L, mEq/L	120	116-124
K	mmol/L, mEq/L	2.8	2.4-3.2
Cl	mmol/L, mEq/L	78	73-83
pH		7.174	7.124-7.224
PCO ₂	mm Hg	59.6	51.6-67.6
	kPa	7.94	6.87-9.00
Glucose/Glu	mg/dL	41	32-50
	g/L	0.41	0.32-0.50
BUN	mmol/L	2.3	1.9-2.7
	mg/dL	53	46-60
Urea	mmol/L	19.0	16.5-21.5
	mg/dL	114	99-129
	g/L	1.14	0.99-1.29


IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022234-01 01
722234-01B
30-Oct-2009

 Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

