

i-STAT Calibration Verification Set B0807519

LOT L1 B08240, L2 B08247, L3 B08255, L4 B08241, L5 B08239

Exp.: 2010-02

CLEW: A18/C18/H18/J18

i-STAT Calibration Verification Set B0807519

LOT L1 B08240, L2 B08247, L3 B08255, L4 B08241, L5 B08239

Exp.: 2010-02

CLEW: A18/C18/H18/J18

CG4+	LOT					J or W
	1	2	3	4	5	
pH	6.826	7.175	7.454	7.684	8.123	
	6.777-6.875	7.126-7.224	7.405-7.503	7.635-7.733	8.074-8.172	
PCO ₂	mmHg	86.87	59.72	25.33	19.08	16.55
		78.36-95.38	53.60-65.84	19.94-30.72	13.94-24.22	11.65-21.45
	kPa	11.580	7.961	3.376	2.543	2.206
		10.445-12.714	7.144-8.777	2.658-4.095	1.858-3.229	1.553-2.859
PO ₂	mmHg	70.1	81.1	108.8	136.7	326.0
		57.1-83.1	68.1-94.1	94.7-122.9	119.0-154.4	274.2-377.8
	kPa	9.34	10.81	14.50	18.22	43.46
		7.62-11.07	9.08-12.54	12.62-16.38	15.86-20.58	36.54-50.37
Lactate/ Lac	mmol/L	17.740	7.884	2.104	0.830	0.442
		15.13-20.35	7.01-8.75	1.70-2.51	0.42-1.24	0.03-0.85
	mg/dL	159.8	71.0	18.96	7.48	3.98
		136.3-183.3	63.2-78.9	15.3-22.6	3.8-11.2	0.3-7.7
	g/L	1.598	0.710	0.1896	0.0748	0.0398
		1.363-1.833	0.632-0.789	0.153-0.226	0.038-0.112	0.003-0.077

CG8+, EG7+, EG6+, G3+, Crea & CG4+	LOT					A, B, C, D, E, F, K, M, P, R, S or T
	1	2	3	4	5	
Na	mmol/L	100.7	120.4	135.4	161.7	180.3
	mEq/L	97.3-104.1	117.0-123.8	131.5-139.3	157.3-166.1	175.1-185.5
K	mmol/L	2.07	2.86	3.71	6.28	7.85
	mEq/L	1.81-2.33	2.60-3.12	3.42-4.00	5.89-6.67	7.35-8.35
pH		6.833	7.174	7.448	7.671	8.093
		6.784-6.882	7.125-7.223	7.399-7.497	7.622-7.720	8.044-8.142
PCO ₂	mmHg	89.00	60.28	26.06	20.41	18.46
		80.28-97.72	54.16-66.40	20.67-31.45	15.27-25.55	13.56-23.36
	kPa	11.864	8.035	3.474	2.721	2.461
		10.701-13.026	7.219-8.852	2.755-4.192	2.035-3.406	1.808-3.114
PO ₂	mmHg	74.7	87.4	117.3	147.3	353.5
		61.7-87.7	74.4-100.4	102.1-132.5	128.2-166.4	301.7-405.3
	kPa	9.96	11.65	15.64	19.64	47.12
		8.23-11.69	9.92-13.38	13.61-17.66	17.09-22.18	40.21-54.03
iCa	mmol/L	2.181	1.450	1.201	0.764	0.323
		2.051-2.311	1.360-1.540	1.126-1.276	0.699-0.829	0.266-0.380
	mg/dL	8.72	5.80	4.80	3.06	1.29
		8.21-9.24	5.44-6.16	4.51-5.10	2.80-3.32	1.06-1.52
Glucose/ Glu	mEq/L	4.36	2.90	2.40	1.53	0.65
		4.10-4.62	2.72-3.08	2.25-2.55	1.40-1.66	0.53-0.76
	mg/dL	28.5	41.3	122.4	281.0	600.5
		19.3-37.7	34.5-48.1	109.9-134.9	252.3-309.7	532.3-668.7
g/L		0.285	0.413	1.224	2.810	6.005
		0.193-0.377	0.345-0.481	1.099-1.349	2.523-3.097	5.323-6.687
	mmol/L	1.58	2.29	6.79	15.60	33.33
		1.07-2.09	1.91-2.67	6.10-7.49	14.00-17.19	29.55-37.12
Creatinine/ Crea	mg/dL	14.81	4.30	1.84	0.64	0.44
		12.16-17.46	3.76-4.84	1.48-2.20	0.28-1.00	0.08-0.80
	umol/L	1309.2	380.1	162.7	56.6	38.9
		1074.8-1543.6	332.5-427.8	131.0-194.3	24.9-88.2	7.2-70.6
Lactate/ Lac	mmol/L	17.888	7.940	2.168	0.847	0.446
		15.26-20.52	7.06-8.82	1.76-2.58	0.44-1.26	0.04-0.85
	mg/dL	161.2	71.5	19.53	7.63	4.02
		137.5-184.9	63.7-79.4	15.9-23.2	4.0-11.3	0.3-7.7
g/L		1.612	0.715	0.1953	0.0763	0.0402
		1.375-1.849	0.637-0.794	0.159-0.232	0.040-0.113	0.003-0.077



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

022237-01 01
722237-01B
30-Oct-2009



i-STAT Calibration Verification Set B0807519

LOT L1 B08240, L2 B08247, L3 B08255, L4 B08241, L5 B08239

Exp.: 2010-02

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+, E3+ & G		LOT			A, B, C, D, E or F	
		1	2	3	4	5
Na	mmol/L	101.6	119.7	133.9	159.1	176.7
	mEq/L	98.2-105.0	116.3-123.1	130.0-137.8	154.7-163.5	171.5-181.9
K	mmol/L	2.08	2.84	3.66	6.17	7.66
	mEq/L	1.82-2.34	2.58-3.10	3.37-3.95	5.78-6.56	7.16-8.16
Cl	mmol/L	65.8	78.0	92.2	115.0	125.8
	mEq/L	61.8-69.8	73.7-82.3	87.7-96.7	109.0-121.0	118.3-133.3
pH		6.832	7.178	7.450	7.686	8.127
		6.783-6.881	7.129-7.227	7.401-7.499	7.637-7.735	8.078-8.176
PCO₂	mmHg	88.53	59.63	24.09	17.64	14.59
		79.86-97.20	53.51-65.75	18.70-29.48	12.50-22.78	9.69-19.49
	kPa	11.801	7.949	3.211	2.351	1.945
		10.645-12.957	7.132-8.765	2.493-3.930	1.666-3.037	1.292-2.598
Glucose/ Glu	mg/dL	28.5	41.3	122.4	281.0	600.5
		19.3-37.7	34.5-48.1	109.9-134.9	252.3-309.7	532.3-668.7
	g/L	0.285	0.413	1.224	2.810	6.005
		0.193-0.377	0.345-0.481	1.099-1.349	2.523-3.097	5.323-6.687
	mmol/L	1.58	2.29	6.79	15.60	33.33
	1.07-2.09	1.91-2.67	6.10-7.49	14.00-17.19	29.55-37.12	
BUN	mg/dL	133.2	53.3	10.1	5.3	2.7
		117.5-148.9	47.0-59.6	7.5-12.7	3.2-7.4	0.6-4.8
Urea	mmol/L	47.55	19.03	3.61	1.89	0.96
		41.93-53.17	16.78-21.28	2.67-4.54	1.14-2.64	0.21-1.71
	mg/dL	285.0	114.1	21.6	11.3	5.8
		251.4-318.7	100.6-127.5	16.0-27.2	6.8-15.8	1.3-10.3
	g/L	2.850	1.141	0.216	0.113	0.058
	2.514-3.187	1.006-1.275	0.160-0.272	0.068-0.158	0.013-0.103	

IVD



CE



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022237-01 01
722237-01B
30-Oct-2009

Abbott