


i-STAT Level 2 Control

LOT B08255

 Exp.: 2010-03


CLEW: A18/C18/H18/J18

CG4+		LOT	J or W	
		\bar{x}	(Mean)	R (Range)
pH			7.454	7.404-7.504
<i>PCO₂</i>	mm Hg		25.3	18.3-32.3
	kPa		3.38	2.44-4.31
<i>PO₂</i>	mm Hg		109	94-124
	kPa		14.5	12.5-16.5
Lactate/Lac	mmol/L		2.10	1.60-2.60
	mg/dL		19.0	14.5-23.5
	g/L		0.190	0.145-0.235

CG8+, EG7+, EG6+, G3+, Crea & CG4+		LOT	A, B, C, D, E, F, K, M, P, R, S or T	
		\bar{x}	(Mean)	R (Range)
Na	mmol/L, mEq/L		135	131-139
K	mmol/L, mEq/L		3.7	3.3-4.1
<i>i</i> Ca	mmol/L		1.20	1.12-1.28
	mg/dL		4.8	4.5-5.1
	mEq/L		2.4	2.2-2.6
pH			7.448	7.398-7.498
<i>PCO₂</i>	mm Hg		26.1	19.1-33.1
	kPa		3.47	2.54-4.41
<i>PO₂</i>	mm Hg		117	102-132
	kPa		15.6	13.6-17.6
Glucose/Glu	mg/dL		122	108-136
	g/L		1.22	1.08-1.36
Creatinine/Crea	mmol/L		6.8	6.0-7.6
	mg/dL		1.8	1.4-2.2
	umol/L		163	127-198
Lactate/Lac	mmol/L		2.17	1.67-2.67
	mg/dL		19.5	15.0-24.0
	g/L		0.195	0.150-0.240

IVD



 Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022239-01 01
722239-01B
30-Oct-2009

 **Abbott**

i-STAT Level 2 Control

LOT B08255

Exp.: 2010-03

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+, E3+ & G		LOT	A, B, C, D, E or F
		\bar{x} (Mean)	R (Range)
Na	mmol/L, mEq/L	134	130-138
K	mmol/L, mEq/L	3.7	3.3-4.1
Cl	mmol/L, mEq/L	92	87-97
pH		7.450	7.400-7.500
PCO ₂	mm Hg	24.1	17.1-31.1
	kPa	3.21	2.28-4.14
Glucose/Glu	mg/dL	122	108-136
	g/L	1.22	1.08-1.36
	mmol/L	6.8	6.0-7.6
BUN	mg/dL	10	7-13
Urea	mmol/L	3.6	2.5-4.7
	mg/dL	22	15-28
	g/L	0.22	0.15-0.28

IVD



CE

Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022239-01 01
722239-01B
30-Oct-2009

