


i-STAT Level 3 Control

LOT B08241

 Exp.: 2010-02


CLEW: A18/C18/H18/J18

CG4+		LOT	J or W	
		\bar{x}	(Mean)	R (Range)
pH			7.684	7.634-7.734
PCO₂	mm Hg		19.1	13.1-25.1
	kPa		2.54	1.74-3.34
PO₂	mm Hg		137	119-155
	kPa		18.2	15.8-20.6
Lactate/Lac	mmol/L		0.83	0.33-1.33
	mg/dL		7.5	3.0-12.0
	g/L		0.075	0.030-0.120

CG8+, EG7+, EG6+, G3+, Crea & CG4+		LOT	A, B, C, D, E, F, K, M, P, R, S or T	
		\bar{x}	(Mean)	R (Range)
Na	mmol/L, mEq/L		162	157-167
K	mmol/L, mEq/L		6.3	5.8-6.8
<i>i</i> Ca	mmol/L		0.76	0.68-0.84
	mg/dL		3.1	2.7-3.4
	mEq/L		1.5	1.4-1.7
pH			7.671	7.621-7.721
PCO₂	mm Hg		20.4	14.4-26.4
	kPa		2.72	1.92-3.52
PO₂	mm Hg		147	129-165
	kPa		19.6	17.2-22.0
Glucose/Glu	mg/dL		281	239-323
	g/L		2.81	2.39-3.23
	mmol/L		15.6	13.3-17.9
Creatinine/Crea	mg/dL		0.6	0.2-1.0
	umol/L		57	21-92
Lactate/Lac	mmol/L		0.85	0.35-1.35
	mg/dL		7.6	3.1-12.1
	g/L		0.076	0.031-0.121

IVD



 Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022240-01 01
722240-01B
30-Oct-2009



i-STAT Level 3 Control

LOT B08241

Exp.: 2010-02

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+, E3+ & G		LOT	A, B, C, D, E or F
		\bar{x} (Mean)	R (Range)
Na	mmol/L, mEq/L	159	154-164
K	mmol/L, mEq/L	6.2	5.7-6.7
Cl	mmol/L, mEq/L	115	109-121
pH		7.686	7.636-7.736
PCO ₂	mm Hg	17.6	11.6-23.6
	kPa	2.35	1.55-3.15
Glucose/Glu	mg/dL	281	239-323
	g/L	2.81	2.39-3.23
	mmol/L	15.6	13.3-17.9
BUN	mg/dL	5	3-7
Urea	mmol/L	1.9	1.2-2.6
	mg/dL	11	7-16
	g/L	0.11	0.07-0.16

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022240-01 01
722240-01B
30-Oct-2009

