


i-STAT PT Level 2 Control

LOT B09189

 Exp.: 2010-06


CLEW: A18/C18/H18/J18

PT **LOT** M, N, S or T

| | \bar{x} (Mean) | R (Range) |
|----------|------------------|-----------|
| PT - INR | 3.00 | 2.3-3.7 |
| PT - sec | 34.2 | 26.0-42.3 |

IVD



 Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022260-01 01
722260-01A
6-Oct-2009

