

i-STAT Cardiac Markers Calibration Verification Control Set

LOT M09173 (L1 M091731, L2 M091732, L3 M091733)

 Exp.: 2010-05

CLEW: A18

cTnl **LOT** **P**

cTnl	ng/mL	L1	L2	L3
		0.29	1.72	21.25
		0.20-0.38	1.20-2.24	11.69-30.81

cTnl **LOT** **R**

cTnl	ng/mL	L1	L2	L3
		0.33	1.93	24.37
		0.23-0.43	1.35-2.51	13.40-35.34

cTnl **LOT** **U**

cTnl	ng/mL	L1	L2	L3
		0.35	2.02	26.09
		0.25-0.46	1.41-2.63	14.35-37.83

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022261-01 01
722261-01A
6-Oct-2009

