

i-STAT Cardiac Markers Level 2 Control

LOT M091732

 Exp.: 2010-05

CLEW: A18

| cTnl | LOT | P | \bar{x} (Mean) | R (Range) |
|-------------|------------|----------|-------------------------|------------------|
| cTnl | | | 1.72 | 1.20-2.24 |

| cTnl | LOT | R | \bar{x} (Mean) | R (Range) |
|-------------|------------|----------|-------------------------|------------------|
| cTnl | | | 1.93 | 1.35-2.51 |

| cTnl | LOT | U | \bar{x} (Mean) | R (Range) |
|-------------|------------|----------|-------------------------|------------------|
| cTnl | | | 2.02 | 1.41-2.63 |

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022263-01 01
722263-01A
6-Oct-2009

