

i-STAT Cardiac Markers Calibration Verification Control Set

LOT M09153 (L1 M091531, L2 M091532, L3 M091533)

 Exp.: 2010-05

CLEW: A18

cTnl **LOT** **P**

cTnl	ng/mL	L1	L2	L3
		0.39	2.05	33.50
		0.27-0.51	1.44-2.67	18.43-48.58

cTnl **LOT** **R**

cTnl	ng/mL	L1	L2	L3
		0.42	2.28	37.00
		0.29-0.55	1.60-2.96	20.35-53.65

cTnl **LOT** **U**

cTnl	ng/mL	L1	L2	L3
		0.49	2.59	42.80
		0.34-0.64	1.81-3.37	23.54-62.06

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022265-01 01
722265-01A
6-Oct-2009

