

CLINIQA Liquid QC™ Cardiac Marker Control for i-STAT

**LOT** Tri-Level 0902274

L1 0901260B (Vial)

L2 0901261B (Vial)

L3 0901262B (Vial)

 Exp.: 2011-02

**CLEW: A18**

**CK-MB** **LOT** **C**

---

CK-MB	ng/mL	L1	L2	L3
		2.8	10.3	54.0
		1.5-4.1	5.7-14.9	29.7-78.3

**IVD**



Manufacturer:  
**Cliniqa Corporation**  
288 Distribution Street  
San Marcos, CA 92078

**EC REP**

CEpartner4U  
Esdoornlaan 13  
3951 DB Maarn  
The Netherlands  
P: (31) 343 442 524  
F: (31) 343 442 162

022273-01 01  
722273-01A  
6-Oct-2009