

CLINIQA Liquid QC™ Cardiac Marker Control for i-STAT

Level 1

LOT 0907174B (Vial)

 Exp.: 2011-09

CLEW: A18

CK-MB **LOT** **C**

		\bar{x} (Mean)	R (Range)
CK-MB	ng/mL	3.2	1.8-4.6

IVD



CE



Manufacturer:
Cliniqa Corporation
288 Distribution Street
San Marcos, CA 92078

EC REP

CEpartner4U
Esdoornlaan 13
3951 DB Maarn
The Netherlands
P: (31) 343 442 524
F: (31) 343 442 162

022278-01 01
722278-01B
16-Dec-2009