

i-STAT BNP Calibration Verification Control Set
LOT M09252 (L1 M092521, L2 M092522, L3 M092523)
Exp.: 2010-02

CLEW: A18

BNP **LOT** **B**

		L1	L2	L3
BNP	pg/mL	99 54-144	945 605-1285	2665 1706-3624

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022285-01 01
722285-01A
6-Oct-2009



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

