


Hematocrit Calibration Verification Set B0908021

**LOT** L1 B09180A, L2 B09168, L3 B09170, L4 B09272, L5 B09271

 Exp.: 2010-11


**CLEW: A18/C18**

**CG8+ , EG7+ & EG6+** **LOT** **A, B, C, D, E, F, M, P, R, S or T**

			1	2	3	4	5
HCT	%	K <sub>3</sub> EDTA	11.8	18.1	34.2	55.9	70.8
			8.8-14.8	15.1-21.1	31.2-37.2	52.5-59.3	66.6-75.0
		K <sub>2</sub> EDTA	12.3	18.9	35.7	58.3	73.8
			9.2-15.4	15.7-22.0	32.5-38.8	54.8-61.8	69.4-78.2

Hematocrit Calibration Verification Set B0908021

**LOT** L1 B09180A, L2 B09168, L3 B09170, L4 B09272, L5 B09271

 Exp.: 2010-11


**CLEW: A18**

**CHEM8+** **LOT** **A, B, E, F, G or H**

			1	2	3	4	5
HCT	%	K <sub>3</sub> EDTA	12.5	19.0	35.1	54.4	70.1
			9.5-15.5	16.0-22.0	32.1-38.1	51.1-57.7	65.9-74.3
		K <sub>2</sub> EDTA	13.0	19.8	36.6	56.7	73.1
			9.9-16.2	16.7-22.9	33.5-39.7	53.3-60.1	68.7-77.5

Hematocrit Calibration Verification Set B0908021

**LOT** L1 B09180A, L2 B09168, L3 B09170, L4 B09272, L5 B09271


 Exp.: 2010-11

**CLEW: A18/C18**

**EC8+, 6+, EC4+ and E3+** **LOT** **A, B, C, D, E or F**

			1	2	3	4	5
HCT	%	K <sub>3</sub> EDTA	12.3	18.7	36.4	54.5	69.4
			9.3-15.3	15.7-21.7	33.4-39.4	51.2-57.8	65.2-73.6
		K <sub>2</sub> EDTA	12.8	19.5	37.9	56.8	72.3
			9.7-16.0	16.4-22.6	34.8-41.1	53.4-60.2	68.0-76.7



 Manufacturer:  
RNA Medical  
Division of Bionostics, Inc.  
Devens, MA 01434



EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

022293-01 01  
722293-01B  
11-Jan-2010