

Hematocrit Calibration Verification Set B0908070

LOT L1 B09180A, L2 B09168, L3 B09170, L4 B09187, L5 B09180B

Exp.: 2010-11

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+ and E3+

LOT

A, B, C, D, E or F

| | | | 1 | 2 | 3 | 4 | 5 |
|------------|---------------------|---------------------|-----------|-----------|-----------|-----------|-----------|
| HCT | % | K ₃ EDTA | 12.3 | 18.7 | 36.4 | 59.1 | 70.1 |
| | | | 9.3-15.3 | 15.7-21.7 | 33.4-39.4 | 55.6-62.6 | 65.9-74.3 |
| | K ₂ EDTA | 12.8 | 19.5 | 37.9 | 61.6 | 73.1 | |
| | | 9.7-16.0 | 16.4-22.6 | 34.8-41.1 | 57.9-65.3 | 68.7-77.5 | |

Hematocrit Calibration Verification Set B0908070

LOT L1 B09180A, L2 B09168, L3 B09170, L4 B09187, L5 B09180B

Exp.: 2010-11

CLEW: A18

CHEM8+

LOT

P or S

| | | | 1 | 2 | 3 | 4 | 5 |
|------------|---------------------|---------------------|-----------|-----------|-----------|-----------|-----------|
| HCT | % | K ₃ EDTA | 12.4 | 18.8 | 34.8 | 55.7 | 67.3 |
| | | | 9.4-15.4 | 15.8-21.8 | 31.8-37.8 | 52.4-59.0 | 63.3-71.3 |
| | K ₂ EDTA | 12.9 | 19.6 | 36.3 | 58.1 | 70.2 | |
| | | 9.8-16.1 | 16.5-22.7 | 33.2-39.4 | 54.6-61.6 | 66.0-74.4 | |

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299



Manufacturer:
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

022297-01 01
722297-01A
16-Oct-2009