

Hematocrit Level 3 Control

LOT B09187

Exp.: 2011-06

**CLEW: A18/C18/H18/J18**

EC8+, 6+, EC4+ and E3+			LOT	A, B, C, D, E or F	
			$\bar{x}$	(Mean)	R (Range)
HCT	K <sub>3</sub> EDTA	%		59.1	55.6-62.6
	K <sub>2</sub> EDTA			61.6	57.9-65.3

Hematocrit Level 3 Control

LOT B09187


Exp.: 2011-06

**CLEW: A18**

CHEM8+			LOT	P or S	
			$\bar{x}$	(Mean)	R (Range)
HCT	K <sub>3</sub> EDTA	%		55.7	52.4-59.0
	K <sub>2</sub> EDTA			58.1	54.6-61.6

IVD



 Manufacturer:  
RNA Medical  
Division of Bionostics, Inc.  
Devens, MA 01434

EC REP

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

022300-01 01  
722300-01A  
16-Oct-2009