


Hematocrit Calibration Verification Set B0958012

LOT L1 B09055, L2 B09048, L3 B09050, L4 B09051, L5 B09062

 Exp.: 2010-07

CLEW: A18/C18/H18/J18


EC8+, 6+, EC4+ and E3+

LOT A, B, C, D, E or F

			1	2	3	4	5
HCT	%	K ₃ EDTA	11.9	18.6	36.3	59.0	69.7
			8.9-14.9	15.6-21.6	33.3-39.3	55.5-62.5	65.5-73.9
	K ₂ EDTA	12.4	19.4	37.8	61.5	72.7	
		9.3-15.5	16.3-22.5	34.7-41.0	57.8-65.2	68.3-77.0	

Hematocrit Calibration Verification Set B0958012

LOT L1 B09055, L2 B09048, L3 B09050, L4 B09051, L5 B09062

 Exp.: 2010-07

CLEW: A18

CHEM8+

LOT

P or S


			1	2	3	4	5
HCT	%	K ₃ EDTA	12.4	18.4	34.4	56.0	67.2
			9.4-15.4	15.4-21.4	31.4-37.4	52.6-59.4	63.2-71.2
	K ₂ EDTA	12.9	19.2	35.9	58.4	70.1	
		9.8-16.1	16.1-22.3	32.7-39.0	54.9-61.9	65.9-74.3	

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

 Manufacturer:
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

022301-01 01
722301-01A
16-Oct-2009