


Hematocrit Calibration Verification Set B0800445

LOT L1 B08340A, L2 B08340B, L3 B08344A, L4 B08344B, L5 B08345

 Exp.: 2010-06

CLEW: A18/C18/H18/J18

EC8+, 6+, EC4+ and E3+

LOT

A, B, C, D, E or F

			1	2	3	4	5
HCT	%	K ₃ EDTA	12.0	18.7	36.4	58.8	69.9
			9.0-15.0	15.7-21.7	33.4-39.4	55.3-62.3	65.7-74.1
	K ₂ EDTA	12.5	19.5	37.9	61.3	72.9	
		9.4-15.6	16.4-22.6	34.8-41.1	57.6-65.0	68.5-77.2	

Hematocrit Calibration Verification Set B0800445

LOT L1 B08340A, L2 B08340B, L3 B08344A, L4 B08344B, L5 B08345

 Exp.: 2010-06

CLEW: A18

CHEM8+

LOT

P or S


			1	2	3	4	5
HCT	%	K ₃ EDTA	12.4	18.4	34.3	55.7	66.9
			9.4-15.4	15.4-21.4	31.3-37.3	52.4-59.0	62.9-70.9
	K ₂ EDTA	12.9	19.2	35.8	58.1	69.7	
		9.8-16.1	16.1-22.3	32.6-38.9	54.6-61.6	65.6-73.9	

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

 **Manufacturer:**
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

022305-01 01
722305-01A
16-Oct-2009