

i-STAT Cardiac Markers Calibration Verification Control Set

LOT M09259 (L1 M092591, L2 M092592, L3 M092593)

 Exp.: 2010-08

CLEW: A18

cTnl **LOT** **R**

		L1	L2	L3
cTnl	ng/mL	0.41	2.11	30.40
		0.29-0.53	1.48-2.74	16.72-44.08

cTnl **LOT** **U**

		L1	L2	L3
cTnl	ng/mL	0.45	2.21	34.44
		0.32-0.59	1.55-2.87	18.94-49.94

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022309-01 01
722309-01A
17-Nov-2009

