

i-STAT BNP Calibration Verification Control Set
LOT M09286 (L1 M092861, L2 M092862, L3 M092863)
Exp.: 2010-04

CLEW: A18

BNP **LOT** **B**

		L1	L2	L3
BNP	pg/mL	102 56-148	1113 712-1514	2962 1896-4028

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022317-01 01
722317-01A
17-Nov-2009



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

