


i-STAT Level 3 Control

LOT B09167

 Exp.: 2010-11

CLEW: A18/C18

CG8+, EG7+, EG6+, G3+, Crea & CG4+		LOT	A, B, C, D, E, F, K, M, P, R, S or T	
		\bar{x}	(Mean)	R (Range)
Na	mmol/L, mEq/L	162		157-167
K	mmol/L, mEq/L	6.2		5.7-6.7
iCa	mmol/L	0.77		0.69-0.85
	mg/dL	3.1		2.8-3.4
	mEq/L	1.5		1.4-1.7
pH		7.684		7.634-7.734
PCO ₂	mm Hg	19.7		13.7-25.7
	kPa	2.63		1.83-3.43
PO ₂	mm Hg	147		129-165
	kPa	19.6		17.2-22.0
Glucose/Glu	mg/dL	283		241-325
	g/L	2.83		2.41-3.25
	mmol/L	15.7		13.4-18.0
Creatinine/Crea	mg/dL	0.7		0.3-1.1
	umol/L	57		22-93
Lactate/Lac	mmol/L	0.84		0.34-1.34
	mg/dL	7.6		3.1-12.1
	g/L	0.076		0.031-0.121

EC8+, 6+, EC4+, E3+ & G		LOT	A, B, C, D, E or F	
		\bar{x}	(Mean)	R (Range)
Na	mmol/L, mEq/L	159		154-164
K	mmol/L, mEq/L	6.2		5.7-6.7
Cl	mmol/L, mEq/L	115		109-121
pH		7.700		7.650-7.750
PCO ₂	mm Hg	17.4		11.4-23.4
	kPa	2.31		1.51-3.11
Glucose/Glu	mg/dL	283		241-325
	g/L	2.83		2.41-3.25
	mmol/L	15.7		13.4-18.0
BUN	mg/dL	5		3-7
Urea	mmol/L	1.8		1.1-2.5
	mg/dL	11		7-15
	g/L	0.11		0.07-0.15

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

022333-01 01
722333-01A
7-Dec-2009

