

Hematocrit Calibration Verification Set B0902431

**LOT** L1 B09320, L2 B09322, L3 B09314, L4 B09323, L5 B09321

**Exp.:** 2011-04

**CLEW: A18/C18**

**CG8+ , EG7+ & EG6+      **LOT**      A, B, C, D, E, F, M, P, R, S or T**

			1	2	3	4	5
<b>HCT</b>	%	K <sub>3</sub> EDTA	12.3	18.1	34.2	54.9	70.3
			9.3-15.3	15.1-21.1	31.2-37.2	51.6-58.2	66.1-74.5
		K <sub>2</sub> EDTA	12.8	18.9	35.7	57.2	73.3
			9.7-16.0	15.7-22.0	32.5-38.8	53.8-60.7	68.9-77.7

Hematocrit Calibration Verification Set B0902431

**LOT** L1 B09320, L2 B09322, L3 B09314, L4 B09323, L5 B09321

**Exp.:** 2011-04

**CLEW: A18/C18**

**EC8+, 6+, EC4+ and E3+      **LOT**      A, B, C, D, E or F**

			1	2	3	4	5
<b>HCT</b>	%	K <sub>3</sub> EDTA	12.5	18.8	34.5	55.7	70.8
			9.5-15.5	15.8-21.8	31.5-37.5	52.4-59.0	66.6-75.0
		K <sub>2</sub> EDTA	13.0	19.6	36.0	58.1	73.8
			9.9-16.2	16.5-22.7	32.8-39.1	54.6-61.6	69.4-78.2

Hematocrit Calibration Verification Set B0902431

**LOT** L1 B09320, L2 B09322, L3 B09314, L4 B09323, L5 B09321

**Exp.:** 2011-04

**CLEW: A18**

**CHEM8+      **LOT**      A, B, E, F, G or H**

			1	2	3	4	5
<b>HCT</b>	%	K <sub>3</sub> EDTA	12.7	18.5	34.5	54.6	69.9
			9.7-15.7	15.5-21.5	31.5-37.5	51.3-57.9	65.7-74.1
		K <sub>2</sub> EDTA	13.2	19.3	36.0	56.9	72.9
			10.1-16.4	16.2-22.4	32.8-39.1	53.5-60.3	68.5-77.2



**Manufacturer:**  
RNA Medical  
Division of Bionostics, Inc.  
Devens, MA 01434



**EMERGO EUROPE**  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

022339-01 01  
722339-01A  
12-Jan-2010