


i-STAT BNP Calibration Verification Control Set

LOT M10067 (L1 M100671, L2 M100672, L3 M100673)

 Exp.: 2010-09

CLEW: A20

BNP **LOT** **B**

| | | L1 | L2 | L3 |
|-----|-------|-----------|-----------|-----------|
| BNP | pg/mL | 122 | 1102 | 2938 |
| | | 67-177 | 705-1499 | 1880-3996 |

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026801-01 01
726801-01A
22-Apr-2010



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

