

i-STAT BNP Calibration Verification Control Set

LOT M10098 (L1 M100981, L2 M100982, L3 M100983)

 Exp.: 2010-10

CLEW: A20

BNP **LOT** **B**

		L1	L2	L3
BNP	pg/mL	127 70-184	1125 720-1530	3310 2118-4502

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026815-01 01
726815-01A
20-May-2010



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

