

i-STAT BNP Calibration Verification Control Set

**LOT** M10131 (L1 M101311, L2 M101312, L3 M101313)

 Exp.: 2010-11

**CLEW: A20**

**BNP** **LOT** **B**

---

		<b>L1</b>	<b>L2</b>	<b>L3</b>
BNP	pg/mL	119 65-173	1135 726-1544	2745 1757-3733

**IVD**



**EC REP**

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026825-01 01  
726825-01A  
2-Jun-2010



Manufacturer:  
Abbott Point of Care Inc.  
Abbott Park, IL  
60064 USA

