


i-STAT CHEM8+ Calibration Verification Level 1B

LOT B09161

 Exp.: 2010-11

CLEW: A20

CHEM8+ **LOT A, B, E, F, G or H**


		1B
TCO ₂	mmol/L	5.2
	mEq/L	0.1-10.3

CHEM8+ **LOT C or D**

		1B
TCO ₂	mmol/L	5.5
	mEq/L	0.4-10.6

IVD



 Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026806-01 01
726806-01A
21-Apr-2010

