


i-STAT CHEM8+ Calibration Verification Level 1B

**LOT** B10099

 Exp.: 2011-09

**CLEW: A20**

**CHEM8+ **LOT** A, B, E, F, G or H**


		<b>1B</b>
TCO <sub>2</sub>	mmol/L	5.4
	mEq/L	0.3-10.5

**CHEM8+ **LOT** C or D**

		<b>1B</b>
TCO <sub>2</sub>	mmol/L	5.4
	mEq/L	0.3-10.5

**IVD**



 **Manufacturer:**  
Abbott Point of Care Inc.  
Abbott Park, IL  
60064 USA

**EC REP**

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026820-01 01  
726820-01A  
2-Jun-2010

