

i-STAT CHEM8+ Calibration Verification Set B0982540

LOT L1 B09168, L2 B09162, L3 B09155, L4 B09174, L5 B09170

Exp.: 2010-11

CLEW: A20

CHEM8+		LOT	A, B, E, F, G or H				
			1	2	3	4	5
Na	mmol/L		100.8	119.2	134.2	160.6	177.9
	mEq/L		97.3-104.3	115.4-123.0	130.2-138.2	156.2-165.0	172.7-183.1
K	mmol/L		2.08	2.84	3.71	6.19	7.77
	mEq/L		1.82-2.34	2.58-3.10	3.42-4.00	5.72-6.66	7.24-8.30
TCO ₂	mmol/L		10.9	17.1	21.8	33.1	54.0
	mEq/L		4.8-17.0	8.7-25.5	15.5-28.1	27.1-39.1	45.3-62.7
iCa	mmol/L		2.205	1.413	1.183	0.743	0.267
			2.063-2.347	1.321-1.505	1.109-1.257	0.678-0.808	0.202-0.332
	mg/dL		8.82	5.65	4.73	2.97	1.07
			8.25-9.39	5.29-6.02	4.44-5.03	2.71-3.23	0.81-1.33
	mEq/L		4.41	2.83	2.37	1.49	0.53
			4.13-4.69	2.64-3.01	2.22-2.51	1.36-1.62	0.40-0.66
Glucose/ Glu	mg/dL		27.6	41.0	121.7	279.0	599.3
			20.4-34.8	34.2-47.8	109.3-134.1	247.3-310.7	531.2-667.4
	g/L		0.276	0.410	1.217	2.790	5.993
			0.204-0.348	0.342-0.478	1.093-1.341	2.473-3.107	5.312-6.674
	mmol/L		1.53	2.28	6.76	15.49	33.26
			1.13-1.93	1.90-2.65	6.06-7.45	13.73-17.25	29.49-37.04
Creatinine/ Crea	mg/dL		14.50	4.14	1.76	0.62	0.36
			11.90-17.10	3.62-4.66	1.40-2.12	0.26-0.98	0.00-0.72
	umol/L		1281.8	366.0	155.6	54.8	31.8
			1052.3-1511.3	320.1-411.8	123.9-187.2	23.2-86.5	0.2-63.5
BUN	mg/dL		143.4	65.9	10.6	5.2	2.1
			126.4-160.4	58.1-73.7	8.0-13.2	3.1-7.3	0.0-4.2
Urea	mmol/L		51.19	23.53	3.78	1.86	0.75
			45.14-57.25	20.75-26.31	2.85-4.72	1.11-2.61	0.00-1.50
	mg/dL		306.9	141.0	22.7	11.1	4.5
			270.6-343.2	124.4-157.7	17.1-28.3	6.6-15.6	0.0-9.0
	g/L		3.069	1.410	0.227	0.111	0.045
			2.706-3.432	1.244-1.577	0.171-0.283	0.066-0.156	0.000-0.090
Cl	mmol/L		75.5	76.4	89.4	109.2	121.7
	mEq/L		71.5-79.5	72.1-80.7	84.9-93.9	103.2-115.2	114.2-129.2

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

i-STAT CHEM8+ Calibration Verification Set B0982540

LOT L1 B09168, L2 B09162, L3 B09155, L4 B09174, L5 B09170

Exp.: 2010-11

CLEW: A20

CHEM8+		LOT	C or D				
			1	2	3	4	5
Na	mmol/L		100.8	119.2	134.2	160.6	177.9
	mEq/L		97.3-104.3	115.4-123.0	130.2-138.2	156.2-165.0	172.7-183.1
K	mmol/L		2.08	2.84	3.71	6.19	7.77
	mEq/L		1.82-2.34	2.58-3.10	3.42-4.00	5.72-6.66	7.24-8.30
TCO ₂	mmol/L		11.6	18.1	23.3	35.5	57.8
	mEq/L		5.5-17.7	9.1-27.1	17.0-29.6	29.5-41.5	48.5-67.1
iCa	mmol/L		2.205	1.413	1.183	0.743	0.267
			2.063-2.347	1.321-1.505	1.109-1.257	0.678-0.808	0.202-0.332
	mg/dL		8.82	5.65	4.73	2.97	1.07
			8.25-9.39	5.29-6.02	4.44-5.03	2.71-3.23	0.81-1.33
	mEq/L		4.41	2.83	2.37	1.49	0.53
			4.13-4.69	2.64-3.01	2.22-2.51	1.36-1.62	0.40-0.66
Glucose/ Glu	mg/dL		25.0	38.7	121.1	282.8	623.4
			17.8-32.2	31.9-45.5	108.7-133.5	250.7-314.9	552.6-694.2
	g/L		0.250	0.387	1.211	2.828	6.234
			0.178-0.322	0.319-0.455	1.087-1.335	2.507-3.149	5.526-6.942
	mmol/L		1.39	2.15	6.72	15.70	34.60
			0.99-1.79	1.77-2.53	6.03-7.41	13.91-17.48	30.67-38.53
Creatinine/ Crea	mg/dL		14.49	4.14	1.76	0.62	0.36
			11.90-17.08	3.62-4.66	1.40-2.12	0.26-0.98	0.00-0.72
	umol/L		1280.9	366.0	155.6	54.8	31.8
			1051.6-1510.3	320.1-411.8	123.9-187.2	23.2-86.5	0.2-63.5
BUN	mg/dL		143.4	65.9	10.6	5.2	2.1
			126.4-160.4	58.1-73.7	8.0-13.2	3.1-7.3	0.0-4.2
Urea	mmol/L		51.19	23.53	3.78	1.86	0.75
			45.14-57.25	20.75-26.31	2.85-4.72	1.11-2.61	0.00-1.50
	mg/dL		306.9	141.0	22.7	11.1	4.5
			270.6-343.2	124.4-157.7	17.1-28.3	6.6-15.6	0.0-9.0
	g/L		3.069	1.410	0.227	0.111	0.045
			2.706-3.432	1.244-1.577	0.171-0.283	0.066-0.156	0.000-0.090
Cl	mmol/L		75.5	76.4	89.4	109.2	121.7
	mEq/L		71.5-79.5	72.1-80.7	84.9-93.9	103.2-115.2	114.2-129.2

026784-01 01
726784-01A
26-Apr-2010

