

i-STAT CHEM8+ Level 2 Control

LOT B09155

Exp.: 2010-11

**CLEW: A20**

CHEM8+	LOT	A, B, E, F, G or H	
		$\bar{x}$ (Mean)	R (Range)
Na	mmol/L, mEq/L	134	130-139
K	mmol/L, mEq/L	3.7	3.4-4.0
iCa	mmol/L	1.18	1.10-1.26
	mg/dL	4.7	4.4-5.1
	mEq/L	2.4	2.2-2.5
TCO <sub>2</sub>	mmol/L, mEq/L	21.8	15.4-28.2
Glucose/Glu	mg/dL	122	105-138
	g/L	1.22	1.05-1.38
	mmol/L	6.8	5.8-7.7
Creatinine/Crea	mg/dL	1.8	1.4-2.2
	umol/L	156	120-191
Cl	mmol/L, mEq/L	89	84-94
BUN	mg/dL	11	8-14
Urea	mmol/L	3.8	2.7-4.9
	mg/dL	23	16-29
	g/L	0.23	0.16-0.29

CHEM8+	LOT	C or D	
		$\bar{x}$ (Mean)	R (Range)
Na	mmol/L, mEq/L	134	130-139
K	mmol/L, mEq/L	3.7	3.4-4.0
iCa	mmol/L	1.18	1.10-1.26
	mg/dL	4.7	4.4-5.1
	mEq/L	2.4	2.2-2.5
TCO <sub>2</sub>	mmol/L, mEq/L	23.3	16.9-29.7
Glucose/Glu	mg/dL	121	105-138
	g/L	1.21	1.05-1.38
	mmol/L	6.7	5.8-7.6
Creatinine/Crea	mg/dL	1.8	1.4-2.2
	umol/L	156	120-191
Cl	mmol/L, mEq/L	89	84-94
BUN	mg/dL	11	8-14
Urea	mmol/L	3.8	2.7-4.9
	mg/dL	23	16-29
	g/L	0.23	0.16-0.29

IVD



Manufacturer:  
Abbott Point of Care Inc.  
Abbott Park, IL  
60064 USA

EC REP

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026786-01 01  
726786-01A  
26-Apr-2010

