

i-STAT Calibration Verification Set B0919245

LOT L1 B09351, L2 B09349, L3 B09342, L4 B09344, L5 B09345

Exp.: 2011-05

CLEW: A20/C20

CG8+, EG7+, EG6+, G3+, Crea & CG4+		A, B, C, D, E, K, M, P, R, S or T				
		1	2	3	4	5
Na	mmol/L	101.3	120.9	136.2	162.4	181.5
	mEq/L	97.8-104.8	117.1-124.7	132.2-140.2	158.0-166.8	176.3-186.7
K	mmol/L	2.08	2.87	3.73	6.30	7.84
	mEq/L	1.82-2.34	2.61-3.13	3.44-4.02	5.83-6.77	7.31-8.37
pH		6.857	7.177	7.450	7.676	8.127
		6.808-6.906	7.128-7.226	7.401-7.499	7.627-7.725	8.078-8.176
PCO ₂	mmHg	91.54	63.93	26.52	19.96	17.64
	kPa	82.57-100.51	57.67-70.19	21.13-31.91	14.82-25.10	12.74-22.54
		12.202	8.522	3.535	2.661	2.351
	11.007-13.398	7.687-9.357	2.817-4.253	1.975-3.346	1.698-3.004	
PO ₂	mmHg	74.1	88.0	116.5	146.3	345.5
	kPa	61.1-87.1	75.0-101.0	101.4-131.6	127.3-165.3	293.7-397.3
		9.88	11.73	15.53	19.50	46.06
	8.15-11.61	10.00-13.46	13.52-17.54	16.97-22.03	39.14-52.97	
iCa	mmol/L	2.214	1.455	1.214	0.782	0.304
	mg/dL	2.072-2.356	1.361-1.549	1.138-1.290	0.717-0.847	0.239-0.369
		8.86	5.82	4.86	3.13	1.22
	mEq/L	8.29-9.42	5.44-6.20	4.55-5.16	2.87-3.39	0.96-1.48
		4.43	2.91	2.43	1.56	0.61
	4.14-4.71	2.72-3.10	2.28-2.58	1.43-1.69	0.48-0.74	
Glucose/ Glu	mg/dL	28.4	41.5	123.2	282.6	602.1
	g/L	21.2-35.6	34.7-48.3	110.6-135.8	250.5-314.7	533.7-670.5
		0.284	0.415	1.232	2.826	6.021
	mmol/L	0.212-0.356	0.347-0.483	1.106-1.358	2.505-3.147	5.337-6.705
		1.58	2.30	6.84	15.69	33.42
	1.17-1.98	1.93-2.68	6.14-7.54	13.90-17.47	29.62-37.22	
Creatinine/ Crea	mg/dL	14.80	4.13	1.77	0.60	0.38
	umol/L	12.15-17.45	3.61-4.65	1.41-2.13	0.24-0.96	0.02-0.74
		1308.3	365.1	156.5	53.0	33.6
	1074.1-1542.6	319.3-410.9	124.8-188.1	21.4-84.7	1.9-65.2	
Lactate/ Lac	mmol/L	17.880	7.940	2.168	0.847	0.446
	mg/dL	15.25-20.51	7.06-8.82	1.76-2.58	0.44-1.26	0.04-0.85
		161.1	71.5	19.53	7.63	4.02
	g/L	137.4-184.8	63.7-79.4	15.9-23.2	4.0-11.3	0.3-7.7
1.611		0.715	0.1953	0.0763	0.0402	
	1.374-1.848	0.637-0.794	0.159-0.232	0.040-0.113	0.003-0.077	

i-STAT Calibration Verification Set B0919245

LOT L1 B09351, L2 B09349, L3 B09342, L4 B09344, L5 B09345

Exp.: 2011-05

CLEW: A20/C20

EC8+, 6+, EC4+, E3+ & G		A, B, C, D or E				
		1	2	3	4	5
Na	mmol/L	101.8	119.8	134.2	159.4	177.0
	mEq/L	98.3-105.3	116.0-123.6	130.2-138.2	155.0-163.8	171.8-182.2
K	mmol/L	2.09	2.85	3.68	6.19	7.66
	mEq/L	1.83-2.35	2.59-3.11	3.39-3.97	5.72-6.66	7.13-8.19
Cl	mmol/L	66.1	78.3	93.0	115.9	126.8
	mEq/L	62.1-70.1	74.0-82.6	88.5-97.5	109.9-121.9	119.3-134.3
pH		6.832	7.167	7.448	7.684	8.150
		6.783-6.881	7.118-7.216	7.399-7.497	7.635-7.733	8.101-8.199
PCO ₂	mmHg	91.67	62.96	25.18	18.26	14.64
	kPa	82.69-100.65	56.79-69.13	19.79-30.57	13.12-23.40	9.74-19.54
		12.220	8.393	3.356	2.434	1.952
	11.022-13.417	7.570-9.215	2.638-4.075	1.748-3.120	1.298-2.605	
Glucose/ Glu	mg/dL	28.4	41.5	123.2	282.6	602.1
	g/L	21.2-35.6	34.7-48.3	110.6-135.8	250.5-314.7	533.7-670.5
		0.284	0.415	1.232	2.826	6.021
	mmol/L	0.212-0.356	0.347-0.483	1.106-1.358	2.505-3.147	5.337-6.705
		1.58	2.30	6.84	15.69	33.42
	1.17-1.98	1.93-2.68	6.14-7.54	13.90-17.47	29.62-37.22	
BUN	mg/dL	130.8	52.4	10.3	5.2	2.1
		115.3-146.3	46.2-58.6	7.7-12.9	3.1-7.3	0.0-4.2
Urea	mmol/L	46.70	18.71	3.68	1.86	0.75
	mg/dL	41.18-52.22	16.50-20.92	2.74-4.61	1.11-2.61	0.00-1.50
		279.9	112.1	22.0	11.1	4.5
	g/L	246.8-313.0	98.9-125.4	16.4-27.7	6.6-15.6	0.0-9.0
		2.799	1.121	0.220	0.111	0.045
	2.468-3.130	0.989-1.254	0.164-0.277	0.066-0.156	0.000-0.090	




Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026776-01 01
726776-01A
26-Apr-2010



i-STAT Calibration Verification Set B0919245

LOT L1 B09351, L2 B09349, L3 B09342, L4 B09344, L5 B09345

Exp.: 2011-05

CLEW: A20/C20

EC8+, 6+, EC4+, E3+ & G		LOT F, G, H, J or K				
		1	2	3	4	5
Na	mmol/L	100.9	120.3	135.4	161.6	180.4
	mEq/L	97.4-104.4	116.5-124.1	131.4-139.4	157.2-166.0	175.2-185.6
K	mmol/L	2.07	2.86	3.70	6.28	7.80
	mEq/L	1.81-2.33	2.60-3.12	3.41-3.99	5.81-6.75	7.27-8.33
Cl	mmol/L	70.1	77.5	90.4	112.0	120.4
	mEq/L	66.1-74.1	73.2-81.8	85.9-94.9	106.0-118.0	112.9-127.9
pH		6.800	7.128	7.403	7.625	8.067
		6.751-6.849	7.079-7.177	7.354-7.452	7.576-7.674	8.018-8.116
PCO₂	mmHg	106.74	75.07	33.25	25.86	23.46
		96.28-117.20	67.71-82.43	27.86-38.64	20.72-31.00	18.56-28.36
	kPa	14.228	10.007	4.432	3.447	3.127
		12.834-15.623	9.026-10.987	3.714-5.151	2.761-4.133	2.474-3.780
Glucose/ Glu	mg/dL	27.1	40.2	121.9	285.0	620.2
		19.9-34.3	33.4-47.0	109.4-134.4	252.6-317.4	549.8-690.6
	g/L	0.271	0.402	1.219	2.850	6.202
		0.199-0.343	0.334-0.470	1.094-1.344	2.526-3.174	5.498-6.906
	mmol/L	1.50	2.23	6.77	15.82	34.42
		1.10-1.91	1.85-2.61	6.07-7.46	14.02-17.62	30.52-38.33
BUN	mg/dL	141.9	63.0	10.8	5.3	2.1
		125.1-158.7	55.6-70.4	8.2-13.4	3.2-7.4	0.0-4.2
Urea	mmol/L	50.66	22.49	3.86	1.89	0.75
		44.67-56.65	19.83-25.15	2.92-4.79	1.14-2.64	0.00-1.50
	mg/dL	303.7	134.8	23.1	11.3	4.5
		267.8-339.6	118.9-150.8	17.5-28.7	6.8-15.8	0.0-9.0
	g/L	3.037	1.348	0.231	0.113	0.045
		2.678-3.396	1.189-1.508	0.175-0.287	0.068-0.158	0.000-0.090



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026776-01 01
726776-01A
26-Apr-2010

