

i-STAT Calibration Verification Set B0953406

LOT L1 B09155, L2 B09153, L3 B09154, L4 B09160, L5 B09156

Exp.: 2010-11

CLEW: A20/C20

CG8+, EG7+, EG6+, G3+, Crea & CG4+		A, B, C, D, E, K, M, P, R, S or T				
		1	2	3	4	5
Na	mmol/L	101.7	120.5	135.9	161.8	181.2
	mEq/L	98.2-105.2	116.7-124.3	131.9-139.9	157.4-166.2	176.0-186.4
K	mmol/L	2.08	2.86	3.71	6.25	7.82
	mEq/L	1.82-2.34	2.60-3.12	3.42-4.00	5.78-6.72	7.29-8.35
pH		6.856	7.179	7.455	7.694	8.115
		6.807-6.905	7.130-7.228	7.406-7.504	7.645-7.743	8.066-8.164
PCO ₂	mmHg	92.90	62.40	26.44	19.09	18.20
	kPa	83.80-102.00	56.28-68.52	21.05-31.83	13.95-24.23	13.30-23.10
		12.384	8.318	3.524	2.545	2.426
PO ₂	mmHg	73.3	86.2	116.1	143.9	348.7
	kPa	60.3-86.3	73.2-99.2	101.1-131.1	125.2-162.6	296.9-400.5
		9.77	11.49	15.48	19.18	46.48
iCa	mmol/L	2.191	1.459	1.209	0.778	0.359
	mg/dL	2.049-2.333	1.364-1.554	1.134-1.284	0.713-0.843	0.294-0.424
		8.76	5.84	4.84	3.11	1.44
mEq/L	8.20-9.33	5.46-6.21	4.54-5.14	2.85-3.37	1.18-1.70	
	4.38	2.92	2.42	1.56	0.72	
	4.10-4.67	2.73-3.11	2.27-2.57	1.43-1.69	0.59-0.85	
Glucose/ Glu	mg/dL	28.4	41.5	123.2	282.6	602.1
	g/L	21.2-35.6	34.7-48.3	110.6-135.8	250.5-314.7	533.7-670.5
		0.284	0.415	1.232	2.826	6.021
mmol/L	0.212-0.356	0.347-0.483	1.106-1.358	2.505-3.147	5.337-6.705	
	1.58	2.30	6.84	15.69	33.42	
	1.17-1.98	1.93-2.68	6.14-7.54	13.90-17.47	29.62-37.22	
Creatinine/ Crea	mg/dL	14.30	4.13	1.77	0.60	0.38
	umol/L	11.74-16.86	3.61-4.65	1.41-2.13	0.24-0.96	0.02-0.74
		1264.1	365.1	156.5	53.0	33.6
Lactate/ Lac	mmol/L	17.880	7.940	2.168	0.847	0.446
	mg/dL	15.25-20.51	7.06-8.82	1.76-2.58	0.44-1.26	0.04-0.85
		161.1	71.5	19.53	7.63	4.02
g/L	137.4-184.8	63.7-79.4	15.9-23.2	4.0-11.3	0.3-7.7	
	1.611	0.715	0.1953	0.0763	0.0402	
	1.374-1.848	0.637-0.794	0.159-0.232	0.040-0.113	0.003-0.077	

i-STAT Calibration Verification Set B0953406

LOT L1 B09155, L2 B09153, L3 B09154, L4 B09160, L5 B09156

Exp.: 2010-11

CLEW: A20/C20

EC8+, 6+, EC4+, E3+ & G		A, B, C, D or E				
		1	2	3	4	5
Na	mmol/L	102.2	119.6	134.0	158.7	176.6
	mEq/L	98.7-105.7	115.8-123.4	130.0-138.0	154.3-163.1	171.4-181.8
K	mmol/L	2.10	2.85	3.67	6.16	7.65
	mEq/L	1.84-2.36	2.59-3.11	3.38-3.96	5.69-6.63	7.12-8.18
Cl	mmol/L	66.2	78.3	92.7	115.3	126.5
	mEq/L	62.2-70.2	74.0-82.6	88.2-97.2	109.3-121.3	119.0-134.0
pH		6.827	7.166	7.452	7.701	8.135
		6.778-6.876	7.117-7.215	7.403-7.501	7.652-7.750	8.086-8.184
PCO ₂	mmHg	92.99	61.42	25.02	17.15	15.07
	kPa	83.88-102.10	55.30-67.54	19.63-30.41	12.01-22.29	10.17-19.97
		12.396	8.187	3.335	2.286	2.009
Glucose/ Glu	mg/dL	28.4	41.5	123.2	282.6	602.1
	g/L	21.2-35.6	34.7-48.3	110.6-135.8	250.5-314.7	533.7-670.5
		0.284	0.415	1.232	2.826	6.021
mmol/L	0.212-0.356	0.347-0.483	1.106-1.358	2.505-3.147	5.337-6.705	
	1.58	2.30	6.84	15.69	33.42	
	1.17-1.98	1.93-2.68	6.14-7.54	13.90-17.47	29.62-37.22	
BUN	mg/dL	131.5	52.8	10.3	5.2	2.1
		116.0-147.0	46.6-59.0	7.7-12.9	3.1-7.3	0.0-4.2
Urea	mmol/L	46.95	18.85	3.68	1.86	0.75
	mg/dL	41.40-52.49	16.62-21.08	2.74-4.61	1.11-2.61	0.00-1.50
		281.4	113.0	22.0	11.1	4.5
g/L	248.1-314.7	99.6-126.3	16.4-27.7	6.6-15.6	0.0-9.0	
	2.814	1.130	0.220	0.111	0.045	
	2.481-3.147	0.996-1.263	0.164-0.277	0.066-0.156	0.000-0.090	



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026772-01 01
726772-01A
26-Apr-2010



i-STAT Calibration Verification Set B0953406

LOT L1 B09155, L2 B09153, L3 B09154, L4 B09160, L5 B09156

Exp.: 2010-11

CLEW: A20/C20

EC8+, 6+, EC4+, E3+ & G		LOT F, G, H, J or K				
		1	2	3	4	5
Na	mmol/L	101.5	120.0	135.3	160.8	180.1
	mEq/L	98.0-105.0	116.2-123.8	131.3-139.3	156.4-165.2	174.9-185.3
K	mmol/L	2.08	2.86	3.69	6.22	7.79
	mEq/L	1.82-2.34	2.60-3.12	3.40-3.98	5.75-6.69	7.26-8.32
Cl	mmol/L	69.9	78.1	89.9	111.3	120.4
	mEq/L	65.9-73.9	73.8-82.4	85.4-94.4	105.3-117.3	112.9-127.9
pH		6.798	7.134	7.408	7.644	8.057
		6.749-6.847	7.085-7.183	7.359-7.457	7.595-7.693	8.008-8.106
PCO₂	mmHg	107.55	72.87	32.88	24.62	23.68
		97.01-118.09	65.73-80.01	27.49-38.27	19.48-29.76	18.78-28.58
	kPa	14.336	9.714	4.383	3.282	3.157
		12.932-15.741	8.762-10.665	3.665-5.101	2.596-3.968	2.504-3.810
Glucose/ Glu	mg/dL	27.1	40.2	121.9	285.0	620.2
		19.9-34.3	33.4-47.0	109.4-134.4	252.6-317.4	549.8-690.6
	g/L	0.271	0.402	1.219	2.850	6.202
		0.199-0.343	0.334-0.470	1.094-1.344	2.526-3.174	5.498-6.906
	mmol/L	1.50	2.23	6.77	15.82	34.42
		1.10-1.91	1.85-2.61	6.07-7.46	14.02-17.62	30.52-38.33
BUN	mg/dL	141.9	62.7	10.6	5.2	2.1
		125.1-158.7	55.3-70.1	8.0-13.2	3.1-7.3	0.0-4.2
Urea	mmol/L	50.66	22.38	3.78	1.86	0.75
		44.67-56.65	19.74-25.03	2.85-4.72	1.11-2.61	0.00-1.50
	mg/dL	303.7	134.2	22.7	11.1	4.5
		267.8-339.6	118.3-150.0	17.1-28.3	6.6-15.6	0.0-9.0
	g/L	3.037	1.342	0.227	0.111	0.045
	2.678-3.396	1.183-1.500	0.171-0.283	0.066-0.156	0.000-0.090	



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026772-01 01
726772-01A
26-Apr-2010

