

i-STAT Cardiac Markers Calibration Verification Control Set

**LOT** M09259 (L1 M092591, L2 M092592, L3 M092593)

**Exp.:** 2010-08

**CLEW: A20**

**cTnl** **LOT** **R**

---

cTnl	ng/mL	<b>L1</b>	<b>L2</b>	<b>L3</b>
		0.41	2.11	30.40
		0.29-0.53	1.48-2.74	16.72-44.08

**cTnl** **LOT** **S**

---

cTnl	ng/mL	<b>L1</b>	<b>L2</b>	<b>L3</b>
		0.39	1.99	34.34
		0.27-0.51	1.39-2.59	18.89-49.79

**cTnl** **LOT** **U**

---

cTnl	ng/mL	<b>L1</b>	<b>L2</b>	<b>L3</b>
		0.45	2.21	34.44
		0.32-0.59	1.55-2.87	18.94-49.94

**IVD**



Manufacturer:  
Abbott Point of Care Inc.  
Abbott Park, IL  
60064 USA

**EC REP**

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026800-01 01  
726800-01A  
22-Apr-2010

