

i-STAT Cardiac Markers Calibration Verification Control Set

LOT M09327 (L1 M093271, L2 M093272, L3 M093273)

 Exp.: 2010-10

CLEW: A20

cTnl **LOT** **R**

cTnl	ng/mL	L1	L2	L3
		0.36	1.87	32.65
		0.25-0.47	1.31-2.43	17.96-47.34

cTnl **LOT** **S**

cTnl	ng/mL	L1	L2	L3
		0.32	1.74	33.95
		0.22-0.42	1.22-2.26	18.67-49.23

cTnl **LOT** **U**

cTnl	ng/mL	L1	L2	L3
		0.37	1.94	34.00
		0.26-0.48	1.36-2.52	18.70-49.30

IVD



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026799-01 01
726799-01A
22-Apr-2010



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

