

i-STAT Cardiac Markers Calibration Verification Control Set

**LOT** M10061 (L1 M100611, L2 M100612, L3 M100613)

 Exp.: 2011-02

**CLEW: A20**

**cTnl** **LOT** **S**

---

|      |       |           |           |             |
|------|-------|-----------|-----------|-------------|
| cTnl | ng/mL | <b>L1</b> | <b>L2</b> | <b>L3</b>   |
|      |       | 0.29      | 1.85      | 23.00       |
|      |       | 0.20-0.38 | 1.30-2.41 | 12.65-33.35 |

**cTnl** **LOT** **U**

---

|      |       |           |           |             |
|------|-------|-----------|-----------|-------------|
| cTnl | ng/mL | <b>L1</b> | <b>L2</b> | <b>L3</b>   |
|      |       | 0.32      | 1.98      | 23.84       |
|      |       | 0.22-0.42 | 1.39-2.57 | 13.11-34.57 |

**IVD**



Manufacturer:  
Abbott Point of Care Inc.  
Abbott Park, IL  
60064 USA

**EC REP**

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026798-01 01  
726798-01A  
22-Apr-2010

