


i-STAT Cardiac Markers Level 2 Control

LOT M092592

 Exp.: 2010-08

CLEW: A20

cTnl **LOT** **R**

		\bar{x} (Mean)	R (Range)
cTnl	ng/mL	2.11	1.48-2.74

cTnl **LOT** **S**

		\bar{x} (Mean)	R (Range)
cTnl	ng/mL	1.99	1.39-2.59

cTnl **LOT** **U**

		\bar{x} (Mean)	R (Range)
cTnl	ng/mL	2.21	1.55-2.87

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026714-01 01
726714-01A
22-Apr-2010

