


i-STAT Cardiac Markers Level 2 Control

LOT M093272

 Exp.: 2010-10

CLEW: A20

| cTnI | LOT | R | \bar{x} (Mean) | R (Range) |
|-------------|------------|----------|------------------------------------|------------------|
| cTnI | | ng/mL | 1.87 | 1.31-2.43 |

| cTnI | LOT | S | \bar{x} (Mean) | R (Range) |
|-------------|------------|----------|------------------------------------|------------------|
| cTnI | | ng/mL | 1.74 | 1.22-2.26 |

| cTnI | LOT | U | \bar{x} (Mean) | R (Range) |
|-------------|------------|----------|------------------------------------|------------------|
| cTnI | | ng/mL | 1.94 | 1.36-2.52 |

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026710-01 01
726710-01A
22-Apr-2010

