


i-STAT Cardiac Markers Level 3 Control

LOT M092593

 Exp.: 2010-08

CLEW: A20

cTnl	LOT	R	\bar{x} (Mean)	R (Range)
cTnl		ng/mL	30.40	16.72-44.08

cTnl	LOT	S	\bar{x} (Mean)	R (Range)
cTnl		ng/mL	34.34	18.89-49.79

cTnl	LOT	U	\bar{x} (Mean)	R (Range)
cTnl		ng/mL	34.44	18.94-49.94

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026715-01 01
726715-01A
22-Apr-2010

