

CLINIQA Liquid QC™ Cardiac Marker Control for i-STAT

LOT Tri-Level 0902274

L1 0901260B (Vial)

L2 0901261B (Vial)

L3 0901262B (Vial)

 Exp.: 2011-02

CLEW: A20

CK-MB **LOT** **C**

CK-MB	ng/mL	L1	L2	L3
		2.6 1.4-3.8	9.8 5.4-14.2	54.2 29.8-78.6

IVD



Manufacturer:
Cliniqa Corporation
288 Distribution Street
San Marcos, CA 92078

EC REP

CEpartner4U
Esdoornlaan 13
3951 DB Maarn
The Netherlands
P: (31) 343 442 524
F: (31) 343 442 162

026804-01 01
726804-01A
22-Apr-2010