


CLINIQA Liquid QC™ Cardiac Marker Control for i-STAT

LOT Tri-Level 0907177

L1 0907174B (Vial)

L2 0907175B (Vial)

L3 0907176B (Vial)

 Exp.: 2011-09

CLEW: A20

CK-MB **LOT** **C**

| | | | | |
|-------|-------|-----------|-----------|-----------|
| CK-MB | ng/mL | L1 | L2 | L3 |
| | | 3.0 | 11.1 | 51.0 |
| | | 1.7-4.4 | 6.1-16.1 | 28.1-74.0 |

IVD



Manufacturer:
Cliniqa Corporation
288 Distribution Street
San Marcos, CA 92078

EC REP

CEpartner4U
Esdoornlaan 13
3951 DB Maarn
The Netherlands
P: (31) 343 442 524
F: (31) 343 442 162

026805-01 01
726805-01A
22-Apr-2010