

Hematocrit Calibration Verification Set B0908070

**LOT** L1 B09180A, L2 B09168, L3 B09170, L4 B09187, L5 B09180B

**Exp.:** 2010-11

**CLEW: A20/C20**

**EC8+, 6+, EC4+ and E3+**

**LOT**

**A, B, C, D, or E**

			1	2	3	4	5
<b>HCT</b>	<b>%</b>	<b>K<sub>3</sub>EDTA</b>	11.8	18.4	36.0	58.9	70.3
			8.8-14.8	15.4-21.4	33.0-39.0	55.4-62.4	66.1-74.5
		<b>K<sub>2</sub>EDTA</b>	12.3	19.2	37.5	61.4	73.3
			9.2-15.4	16.1-22.3	34.4-40.7	57.7-65.1	68.9-77.7

**IVD**



Manufacturer:  
RNA Medical  
Division of Bionostics, Inc.  
Devens, MA 01434

**EC REP**

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026760-01 01  
726760-01A  
26-Apr-2010