


Hematocrit Calibration Verification Set B0958012

LOT L1 B09055, L2 B09048, L3 B09050, L4 B09051, L5 B09062

 Exp.: 2010-07

CLEW: A20/C20

EC8+, 6+, EC4+ and E3+

LOT


A, B, C, D, or E

			1	2	3	4	5
HCT	%	K₃EDTA	11.7	18.3	35.9	59.1	70.3
			8.7-14.7	15.3-21.3	32.9-38.9	55.6-62.6	66.1-74.5
		K₂EDTA	12.2	19.1	37.4	61.6	73.3
			9.1-15.3	16.0-22.2	34.3-40.6	57.9-65.3	68.9-77.7

IVD



EC REP

 **Manufacturer:**
 RNA Medical
 Division of Bionostics, Inc.
 Devens, MA 01434

EMERGO EUROPE
 Molenstraat 15
 2513 BH, The Hague
 The Netherlands
 Tel: +31.70.345.8570
 Fax: +31.70.346.7299

026756-01 01
 726756-01A
 26-Apr-2010