

Hematocrit Level 1 Control

**LOT** B10117

**Exp.:** 2012-03

**CLEW: A20/C20**

CG8+, EG7+ & EG6+			<b>LOT</b>	C, M, P, R, S or T	
			$\bar{x}$	<b>(Mean)</b>	<b>R (Range)</b>
HCT	K <sub>3</sub> EDTA	%		17.4	14.4-20.4
	K <sub>2</sub> EDTA			18.1	15.0-21.3

EC8+, 6+, EC4+ or E3+			<b>LOT</b>	A, B, C, D or E	
			$\bar{x}$	<b>(Mean)</b>	<b>R (Range)</b>
HCT	K <sub>3</sub> EDTA	%		18.5	15.5-21.5
	K <sub>2</sub> EDTA			19.3	16.2-22.4

EC8+, 6+, EC4+ or E3+			<b>LOT</b>	F, G, H, J or K	
			$\bar{x}$	<b>(Mean)</b>	<b>R (Range)</b>
HCT	K <sub>3</sub> EDTA	%		18.2	15.2-21.2
	K <sub>2</sub> EDTA			19.0	15.8-22.1

Hematocrit Level 1 Control

**LOT** B10117

**Exp.:** 2012-03


**CLEW: A20**

CHEM8+			<b>LOT</b>	A, B, E, F, G or H	
			$\bar{x}$	<b>(Mean)</b>	<b>R (Range)</b>
HCT	K <sub>3</sub> EDTA	%		17.6	14.6-20.6
	K <sub>2</sub> EDTA			18.3	15.2-21.5

CHEM8+			<b>LOT</b>	C or D	
			$\bar{x}$	<b>(Mean)</b>	<b>R (Range)</b>
HCT	K <sub>3</sub> EDTA	%		18.7	15.7-21.7
	K <sub>2</sub> EDTA			19.5	16.4-22.6

**IVD**



 Manufacturer:  
RNA Medical  
Division of Bionostics, Inc.  
Devens, MA 01434

**EC REP**

EMERGO EUROPE  
Molenstraat 15  
2513 BH, The Hague  
The Netherlands  
Tel: +31.70.345.8570  
Fax: +31.70.346.7299

026830-01 01  
726830-01A  
24-Jun-2010