

Hematocrit Level 2 Control

LOT B09170

 Exp.: 2011-05

CLEW: A20/C20


CG8+ , EG7+ & EG6+			LOT	C, M, P, R, S or T	
			\bar{x} (Mean)	R (Range)	
HCT	K ₃ EDTA	%	33.2	30.2-36.2	
	K ₂ EDTA		34.6	31.5-37.7	

EC8+, 6+, EC4+ and E3+			LOT	A, B, C, D or E	
			\bar{x} (Mean)	R (Range)	
HCT	K ₃ EDTA	%	36.0	33.0-39.0	
	K ₂ EDTA		37.5	34.4-40.7	

EC8+, 6+, EC4+, E3+ & G			LOT	F, G, H, J or K	
			\bar{x} (Mean)	R (Range)	
HCT	K ₃ EDTA	%	34.3	31.3-37.3	
	K ₂ EDTA		35.8	32.6-38.9	

Hematocrit Level 2 Control

LOT B09170

 Exp.: 2011-05


CLEW: A20

CHEM8+			LOT	A, B, E, F, G or H	
			\bar{x} (Mean)	R (Range)	
HCT	K ₃ EDTA	%	33.9	30.9-36.9	
	K ₂ EDTA		35.3	32.2-38.5	

CHEM8+			LOT	C or D	
			\bar{x} (Mean)	R (Range)	
HCT	K ₃ EDTA	%	33.6	30.6-36.6	
	K ₂ EDTA		35.0	31.9-38.2	

IVD



 Manufacturer:
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026750-01 01
726750-01A
26-Apr-2010