


Hematocrit Level 2 Control

LOT B09314

 Exp.: 2011-10

CLEW: A20/C20


CG8+, EG7+ & EG6+			LOT	C, M, P, R, S or T	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		33.3	30.3-36.3
	K ₂ EDTA			34.7	31.6-37.8

EC8+, 6+, EC4+ and E3+			LOT	A, B, C, D or E	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		36.1	33.1-39.1
	K ₂ EDTA			37.6	34.5-40.8

EC8+, 6+, EC4+, E3+ & G			LOT	F, G, H, J or K	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		34.3	31.3-37.3
	K ₂ EDTA			35.8	32.6-38.9

Hematocrit Level 2 Control

LOT B09314

 Exp.: 2011-10


CLEW: A20

CHEM8+			LOT	A, B, E, F, G or H	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		33.7	30.7-36.7
	K ₂ EDTA			35.1	32.0-38.3

CHEM8+			LOT	C or D	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		33.6	30.6-36.6
	K ₂ EDTA			35.0	31.9-38.2

IVD



 Manufacturer:
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026754-01 01
726754-01A
26-Apr-2010