

Hematocrit Level 3 Control

LOT B09272

Exp.: 2011-08

CLEW: A20/C20

| CG8+, EG7+ & EG6+ | | | LOT | C, M, P, R, S or T | |
|-------------------------|---------------------|---|------------------|--------------------|--|
| | | | \bar{x} (Mean) | R (Range) | |
| HCT | K ₃ EDTA | % | 53.7 | 50.5-56.9 | |
| | K ₂ EDTA | | 56.0 | 52.6-59.3 | |
| EC8+, 6+, EC4+ and E3+ | | | LOT | A, B, C, D or E | |
| | | | \bar{x} (Mean) | R (Range) | |
| HCT | K ₃ EDTA | % | 55.5 | 52.2-58.8 | |
| | K ₂ EDTA | | 57.9 | 54.4-61.3 | |
| EC8+, 6+, EC4+, E3+ & G | | | LOT | F, G, H, J or K | |
| | | | \bar{x} (Mean) | R (Range) | |
| HCT | K ₃ EDTA | % | 54.5 | 51.2-57.8 | |
| | K ₂ EDTA | | 56.8 | 53.4-60.2 | |

Hematocrit Level 3 Control

LOT B09272


Exp.: 2011-08

CLEW: A20

| CHEM8+ | | | LOT | A, B, E, F, G or H | |
|--------|---------------------|---|------------------|--------------------|--|
| | | | \bar{x} (Mean) | R (Range) | |
| HCT | K ₃ EDTA | % | 54.6 | 51.3-57.9 | |
| | K ₂ EDTA | | 56.9 | 53.5-60.3 | |
| CHEM8+ | | | LOT | C or D | |
| | | | \bar{x} (Mean) | R (Range) | |
| HCT | K ₃ EDTA | % | 54.5 | 51.2-57.8 | |
| | K ₂ EDTA | | 56.8 | 53.4-60.2 | |

IVD



 **Manufacturer:**
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026751-01 01
726751-01A
26-Apr-2010