


Hematocrit Level 3 Control

LOT B09323


 Exp.: 2011-10

CLEW: A20/C20

CG8+, EG7+ & EG6+			LOT	C, M, P, R, S or T	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		54.0	50.8-57.2
	K ₂ EDTA			56.3	52.9-59.7
EC8+, 6+, EC4+ and E3+			LOT	A, B, C, D or E	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		55.6	52.3-58.9
	K ₂ EDTA			58.0	54.5-61.4
EC8+, 6+, EC4+, E3+ & G			LOT	F, G, H, J or K	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		54.6	51.3-57.9
	K ₂ EDTA			56.9	53.5-60.3

Hematocrit Level 3 Control

LOT B09323


 Exp.: 2011-10

CLEW: A20

CHEM8+			LOT	A, B, E, F, G or H	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		54.5	51.2-57.8
	K ₂ EDTA			56.8	53.4-60.2
CHEM8+			LOT	C or D	
			\bar{x}	(Mean)	R (Range)
HCT	K ₃ EDTA	%		54.8	51.5-58.1
	K ₂ EDTA			57.1	53.7-60.6

IVD



 Manufacturer:
RNA Medical
Division of Bionostics, Inc.
Devens, MA 01434

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026755-01 01
726755-01A
26-Apr-2010