

i-STAT Level 3 Control

LOT B09160

Exp.: 2010-11

CLEW: A20/C20

CG8+, EG7+, EG6+, G3+, Crea & CG4+		LOT	A, B, C, D, E, K, M, P, R, S or T	
		\bar{x}	(Mean)	R (Range)
Na	mmol/L, mEq/L	162		157-167
K	mmol/L, mEq/L	6.3		5.7-6.8
iCa	mmol/L	0.78		0.70-0.86
	mg/dL	3.1		2.8-3.4
	mEq/L	1.6		1.4-1.7
pH		7.694		7.644-7.744
PCO ₂	mm Hg	19.1		13.1-25.1
	kPa	2.54		1.74-3.34
PO ₂	mm Hg	144		126-162
	kPa	19.2		16.8-21.6
Glucose/Glu	mg/dL	283		241-325
	g/L	2.83		2.41-3.25
	mmol/L	15.7		13.4-18.0
Creatinine/Crea	mg/dL	0.6		0.2-1.0
	umol/L	53		18-88
Lactate/Lac	mmol/L	0.85		0.35-1.35
	mg/dL	7.6		3.1-12.1
	g/L	0.076		0.031-0.121

EC8+, 6+, EC4+, E3+ & G		LOT	A, B, C, D or E	
		\bar{x}	(Mean)	R (Range)
Na	mmol/L, mEq/L	159		154-164
K	mmol/L, mEq/L	6.2		5.6-6.7
Cl	mmol/L, mEq/L	115		109-121
pH		7.701		7.651-7.751
PCO ₂	mm Hg	17.2		11.2-23.2
	kPa	2.29		1.49-3.09
Glucose/Glu	mg/dL	283		241-325
	g/L	2.83		2.41-3.25
	mmol/L	15.7		13.4-18.0
BUN	mg/dL	5		3-7
Urea	mmol/L	1.9		1.1-2.6
	mg/dL	11		7-15
	g/L	0.11		0.07-0.15

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA

EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026775-01 01
726775-01A
26-Apr-2010



i-STAT Level 3 Control

LOT B09160

Exp.: 2010-11

CLEW: A20/C20

EC8+, 6+, EC4+, E3+ & G		LOT	F, G, H, J or K
		\bar{x} (Mean)	R (Range)
Na	mmol/L, mEq/L	161	156-166
K	mmol/L, mEq/L	6.2	5.7-6.8
Cl	mmol/L, mEq/L	111	105-117
pH		7.644	7.594-7.694
PCO ₂	mm Hg	24.6	18.6-30.6
	kPa	3.28	2.48-4.08
Glucose/Glu	mg/dL	285	243-327
	g/L	2.85	2.43-3.27
	mmol/L	15.8	13.5-18.1
BUN	mg/dL	5	3-7
Urea	mmol/L	1.9	1.1-2.6
	mg/dL	11	7-15
	g/L	0.11	0.07-0.15

IVD



Manufacturer:
Abbott Point of Care Inc.
Abbott Park, IL
60064 USA



EC REP

EMERGO EUROPE
Molenstraat 15
2513 BH, The Hague
The Netherlands
Tel: +31.70.345.8570
Fax: +31.70.346.7299

026775-01 01
726775-01A
26-Apr-2010

